



PRODUCT DATA

I-Product Description & Classification		
I/1	Designation	Care Sets
I/2	General Description	Set : Bandage 5
I/3	GMDN N°	40247
I/4	Reference Number	4908017
I/5	Class	<input type="checkbox"/> I <input checked="" type="checkbox"/> Is <input type="checkbox"/> IIa According to Directive 93/42/EEC, Annex IX, Rule 4
I/6	Notify Body	CE ₀₁₂₀
I/7	Composition Raw Material	1 Anatomic Tweezer, 14-14.5cm 6 Woven Compresses, 5x5cm, 12 plies 1 Drape, 50x50cm
I/8	Sterile	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No If, yes <input checked="" type="checkbox"/> ETO <input type="checkbox"/> Gamma <input type="checkbox"/> Steam
I/9	First Packaging	Number of pieces in Packaging: 36 sets per carton
I/10	Picture	

II- Description of use and operating of device		
II/1	Intended use/indication:	Set that contains all the various instruments for the clean or ablation of suture or pose a patch around the wound. Alone, single use and transient use (<60min)
II/2	User:	Medical staff and Professional
II/3	Contraindication of use	No recognized contraindication
II/4	Special requirement for use	n/a
II/5	Handling and storage conditions:	Standard condition or atmospheric condition







III- Supplementary information		
III/1	Compatibility with others medical devices	YES
III/2	Device contents animal substances	NO
III/3	Device contents medicinal substances	NON



Zone industrielle
Route de Pra de Plan 1
CH-1618 CHATEL-ST-DENIS
Switzerland

TEL : +41 (0) 21 948 92 74
FAX : +41 (0) 21 948 90 36
E-mail : info@applimed.ch
Website : www.applimed.ch

III/4	Device comes into contact with substances	YES
III/5	Used with Accessories	NO
III/6	Used in combination with other Medical Device	NO

IV- MANUFACTURING		
IV/1	Manufacturing Operations <input type="checkbox"/> peel pack <input type="checkbox"/> pouches <input type="checkbox"/> boxes <input checked="" type="checkbox"/> cartons <input type="checkbox"/> cleaning <input checked="" type="checkbox"/> sterilization	36 sets per cartons
IV/2	<u>Labelling: According to Directive 93/42/EEC Annex I § 13.3 and ISO 15223-1</u> <ul style="list-style-type: none"> Name and description of product Name and address of manufacturer LOT N° Article N° Year of manufacturing Expiry date Quantity Sterile Method of sterilisation Indication for single use Not use if the pack is opened CE 0120 	Example of a label Set Pansement 5 1 Pincette 13cm, bleue 6 Compresses de gaze 5x5 cm, 12 plis 1 Champ 50x50cm 36 Sets/carton VERBANDSET 5 (Plastik) 1 Pinzette anatomisch (Plastik) 6 Gazekompressen 5x5 cm 12 fach 1 Abdecktuch 50x50cm 36 Sets/Karton Distributor: Kuhn und Bieri AG- 3098 Köniz - www.kuhnbieteri.ch  SV032400 REF 4908017  2019-01 LOT 14.01.13  STERILE EO    Applimed SA Pra de Plan 1 CH-1618 Châtel-St-Denis
IV/3	Place of sterilisation:	ETO gas sterilization according to ISO 11135-1, By CSSR (Centre de sterilisation de Suisse Romande)