



# TR9650 Mobile bath lift system

Hygiene Equipment  
Designed for People



## Declaration of Conformity

<b>Valid from:</b>	2021-04-01
<b>Manufacturer:</b>	TR Equipment AB Rundelgatan 2 573 35 Tranås Sweden
<b>SRN: / Notified body:</b>	N/A
<b>Statement:</b>	This declaration of conformity is issued under the sole responsibility of TR Equipment AB
<b>Basic UDI-DI:</b>	7350016060406J
<b>Product:</b>	TR 9650 – Mobile patient bath lift system
<b>Product variants:</b>	TR9650 Bath Chair lift, TR9650 Bath stretcher lift, TR9650 Bath Combi lift with chair, Stretcher side parts for combi lift with chair.
<b>Presumption of compliance</b>	ISO 10535:2006
<b>Classification:</b>	MDR Class I
<b>MDR Annex:</b>	VIII
<b>Accessories approved by manufacturer:</b>	Lift Weighing Kit MHS-2800 including display and load cells fulfills OIML Class III standard

We hereby declare that above mentioned devices comply with the European Medical Device Regulations MDR 2017/745, Swedish law 1993:584 regarding Medical devices, Regulation 1993:876 regarding Medical devices and regulation LVFS 2003:11 from the Swedish Medical Products Agency.

Tranås 2022-04-28

**TR Equipment AB**

Mattias Franzén, MD

#### Head office

TR Equipment AB, Sweden  
Phone: +46 140 38 50 60  
info@trequipment.se  
www.trequipment.se

#### Group Subsidiaries

TR Group Inc, US, Phone: +1 (800) 752 6900  
TR Equipment Ltd, UK, Phone: +44 (0) 844 335 8386  
TRanspoRehaMed GmbH, Germany, Phone: +49 (0) 212 645850-0



**TR EQUIPMENT**