

## Declaration of Conformity TOPRO Taurus H

Document ID:	<b>6 344</b>	Document type:	<b>Declaration of conformity</b>
Revision:	<b>1</b>	Document owner:	<b>Quality</b>
Prepared date:	<b>11.05.2021</b>	Prepared by:	<b>Cioroiu, Cosmin</b>
Last revised:	<b>10.05.2021</b>	Revised by:	<b>Cioroiu, Cosmin</b>
Approved:	<b>Ja</b>	Approved by:	<b>Amlien, Åsmund</b>
Approved date:	<b>11.05.2021</b>		

Last change written in *italic* and **green** font.

*Published: 11.05.2021*

*This Declaration of Conformity is issued under the sole responsibility of the manufacturer. This DoC is designed according to Annex IV of Medical Device Regulation (EU) 2017/745.*

As Legal Manufacturer, we

TOPRO Industri AS  
Rambekkevegen 5  
NO-2816 Gjøvik  
NORWAY

SRN: NO-MF-000003447

hereby declare under our sole responsibility that the following CE marked device(s)

Product/trade name(s)	TOPRO Taurus H	
Intended Purpose	The device shall give support to users with reduced balance and/or reduced walking ability	
Model Number(s) and Name(s)	814785	TOPRO Taurus H Basic
	814787	TOPRO Taurus H Premium
Variant Number(s)		
Basic UDI-DI	705432TAH1478SV	

are classified per rule 1 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class I devices in accordance with all applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

Document title: **Declaration of Conformity TOPRO  
Taurus H**  
Document ID: **6 344**  
Approved: **Ja**

Document type: **Declaration of conformity**  
Revision: **1**  
Approved by: **Amlien, Åsmund**

The object of the Declaration described above is in conformity with the following regulations and standards:

EU Regulation	Medical Device Regulation (EU) 2017/745
NS-EN ISO 9999:2016	Assistive products for persons with disability - Classification and terminology
NS-EN 12182:2012	Assistive products for persons with disability - General requirements and test methods
NS-EN 1985:1999	Walking aids - General requirements and test methods
NS-EN ISO 11199-3:2005	Walking aids manipulated by both arms. Requirements and test methods - Part 3: Walking tables

TOPRO hereby confirm that all models/variants and their original accessories are produced and tested in accordance the above mentions regulations and standards. All the technical documentation for the device(s) are stored at the manufacturer.

The user manuals are attached with the products.

Gjøvik / 2021.05.10

Åsmund Amlien / QA-manager

## Document history

### Replacements

DATE	HISTORY	REV	SIGN
2021.05.10	Replaces Doc.Id:5292	4	ÅA

### Changes

DATE	HISTORY	REV	SIGN