

**Declaration of Conformity (Certificate Ref. JH011 Rev. C)**

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, importer & distributor –

**Joerns Healthcare Limited**, Drakes Broughton Business Park, Worcester Road, Pershore, Worcestershire, UK. WR10 2AG.

The device that is covered by the present declaration is in conformity with Medical Device Regulation (EU) 2017/745.

**Product Name:** Oxford Classic

**Description:** The Oxford Classic range of hoists are ergonomically designed mobile devices for the safe lifting and handling of patients for the transfer or moving a person over short distances.

**Product Codes:**

OXF-MAJOR	OXF-MIDI	OXF-MINI
OXF-MAJORE-EL-UK	OXF-MIDIE-EL-AU	OXF-MINIE-EL-EU
OXF-MAJORE-EU	OXF-MIDIE-EL-UK	OXF-MINIE-EL-UK
OXF-MAJORE-UK	OXF-MIDIE-EM-UK	OXF-MINIE-EU
	OXF-MIDIE-EU	OXF-MINIE-UK
OXF-MAXIE-UK	OXF-MIDIE-UK	

**Actor ID/SRN:** GB-MF-000013472 (UK Rep)      NL-AR-000006632 (NL Rep)

**Basic UDI-DI (GMN):** 50556184OXF-CLASSICNS    **GMDN:** 12330    **EMDN:** V08050301

**Classification Type:** Class I [Rule 1 of MDR 2017/745, Annex VIII, Chapter III, Section 4.1]

**Authorised Representative:** Joerns Healthcare BV, Zeilschip 26,3991 CT Houten, Netherlands

This device was tested and is in conformity with all relevant clauses within the following standards:

Standard	Description
BS EN ISO 10535:2021	Hoists for the transfer of disabled persons
BS EN ISO 12182:2012	Assistive products for persons with disability. General requirements & test methods
BS EN ISO 14971:2019	Application of Risk Management to medical devices
BS EN ISO 13485:2016	Regulatory requirements for medical device quality management systems

**Issued By:** Stan Snijders – Managing Director of Joerns Healthcare, Europe

**Issue Date:** 01/11/2023

**Location:** UK / Netherlands

**On behalf of:** Joerns Healthcare Limited, UK

**Signature:** 