	Document Name	Declaration of Conformity			Page	Page 1 of 2
	Document No.	CE-UB-02-01	Version	V01	Effective Date	2021-03-11

## EC Declaration of Conformity

According to Art. 19 of Regulation (EU) 2017/745 on Medical Devices

Name: Hangzhou Primecare Medical Co., Ltd.

Manufacturer: Add: Room 408-409, Zancheng Center West, Shangcheng District, 310008 Hangzhou, PEOPLE'S REPUBLIC OF CHINA

SRN: Not available yet  
Name: MedPath GmbH

European Representative: Add: Mies-van-der-Rohe-Strasse 8, 80807 Munich, Germany

SRN: DE-AR-000000087

Product Name: Urine Bag

Trade name: Primecare

Product REF	Primecare Product REF	GHC Product REF
	8101101	04/01/500US/10/GHC
	8101102	04/01/500US/30/GHC
	8101103	04/01/500US/50/GHC
	8101504	04/01/750UF/50/GHC
	8101501	04/01/750US/10/GHC
	8101502	04/01/750US/30/GHC
	8101503	04/01/750US/50/GHC
	8101702	04/06/2000/90/GHC
	8101705	04/06/2000/130/GHC
	8101701	04/07/2000/90/GHC
	8101602	04/07/1500/90/00
	8101601	04/07/1500/90/01
	8101402	04/01/750USE/12/GHC
	8101707	04/CH/2000US/90/GHC
	8101708	04/CH/1500USo/90/GHC
	8101709	04/CH/2000USo/90/GHC


Basic UDI 697236069UB01KM

Intended purpose This product is used for draining accumulated urine.

Classification acc. to MDR Ax. VIII: Class I, Rule 1

Applied Common Specification/standard: EN ISO 13485: 2016; EN ISO 14971: 2019; EN ISO 15223-1: 2016; EN 1041: 2008/A1: 2013; ISO 8669-2:1996

Conformity Assessment Procedure: MDR Annex II + Annex III

	Document Name	Declaration of Conformity			Page	Page 2 of 2
	Document No.	CE-UB-02-01	Version	V01	Effective Date	2021-03-11

We, the manufacturer, herewith declare under our sole responsibility that the above-mentioned products meet the provisions of the Regulation (EU) 2017/745 on Medical Devices (MDR). All supporting documentations are retained under the premises of the manufacturer.

Place of Issue: Hangzhou, CHINA

Date of Issue: 2022-11-16

Signature:



Yan Xueqing

Position: Management Representative